



---

# Premarket Approval & Humanitarian Device Exemption Applications

Quynh Hoang  
Senior Regulatory Consultant  
FDA & Life Sciences  
Government Matters Practice

# Disclaimer

---



The views expressed here are solely mine and not of my firm or any of its clients.



# Outline

---

1. Definitions
2. Premarket Approval Application (PMA)
3. Humanitarian Device Exemption Application (HDE)
4. Breakthrough and Safer Designations for Submissions
5. Key Takeaways
6. Key FDA Resources



# Definitions

---

- CFR ≡ Code of Federal Regulations (21 CFR)
- CRF ≡ Case Report Form
- FDCA ≡ Federal Food Drug & Cosmetic Act
- FY ≡ Fiscal Year is from Oct. 1st to Sept. 30th
- IDE ≡ Investigational Device Exemption application
- IRB ≡ Institutional Review Board
- OCE ≡ Office of Communication and Education
- OPEQ ≡ Office of Product Evaluation and Quality
- OSEL ≡ Office of Science and Engineering Laboratories
- QSR ≡ Quality System Regulation
- Q-Sub ≡ Pre-submission feedback program

# Premarket Approval Application (PMA)

—



Number of Original PMAs and Panel-track Supplements*		
Review Office	FY 2020	FY 2021
OHT1	6	8
OHT2	23	19
OHT3	7	4
OHT4	5	6
OHT5	4	9
OHT6	2	4
OHT7	30	28
<b>Total PMAs</b>	<b>77</b>	<b>78</b>
510(k)s	3700	3985
De Novos	69	63

\*Number Received by CDRH, per Feb. 2022 Performance Report



## Original PMA (when applicable)

---

**Class III device** (FDCA 513(a)(1)(C) and CFR 860.3 (c)(3)):

1. General Controls are insufficient (Not Class I);
2. General and Special Controls are insufficient (Not Class II);  
and,
3. One of the following:
  - a. life-supporting,
  - b. life-sustaining,
  - c. of substantial importance in preventing impairment of human health, or
  - d. if presents a potential unreasonable risk.

# Classification Polling Question - Background

**The Epicor™ Ablation System is intended for ablating cardiac tissue.**

- Physical State: Tip (c) delivers focused energy to ablate the tissue. Tip goes around the beating heart.
- Technical Method: Device delivers high frequency focused ultrasound (HIFU) to heat and create lesion at the target tissue volume.

Epicor™ System Console (a), Epicor™ UltraWand LP (b) and medical UltraCinch LP ablation device (c).  
LP: low profile.



(a)

(b)

(c)

Hindawi.com




# Classification Polling Question

The Epicor™ Ablation System is intended for ablating cardiac tissue.

- Physical State: Tip (c) delivers focused energy to ablate the tissue. Tip goes around the beating heart.
- Technical Method: Device delivers high frequency focused ultrasound (HIFU) to heat and create lesion at the target tissue volume.

Epicor™ Ablation System



Hindawi.com

What is the Class that FDA would likely assign to device?

Please choose one of the following:

1. Class I (general controls are sufficient);
2. Class II (general and special controls are sufficient);
3. Class III (general and PMA controls); or,
4. I need to confirm some details first [please type your question in Q/A].



# Original PMA (contents)

---

## Contents (FDCA 515(c) and CFR 814.20)

1. Summary of safety and effectiveness data (SSED);
2. Indications;
3. Device Description (**reference Master File(s)**);
4. Preclinical data;
- 5. Clinical data (valid scientific evidence);**
6. Labeling;
- 7. Manufacturing;**
- 8. Post-approval study proposal;**
9. ...



## Original PMA (contents)

---

Clinical data-**valid scientific evidence** (FDCA 513(a)(3) and CFR 860.7(c)(2) ):

- well-controlled investigations;
- partially controlled studies (in CFR);
- studies and objective trials without matched controls (in CFR);
- well-documented case histories conducted by qualified experts (in CFR); and/or
- reports of significant human experience with a marketed device (in CFR).

And, data may be from **Outside U.S.** (CFR 814.15).



# Original PMA (submission options)

---

## – Modular

1. shell,
2. device description,
3. bench testing,
4. manufacturing info,
5. ..., and
6. clinical and labeling.

## – Traditional

### Contents of an Original PMA

1. Indications;
2. Device Description;
3. Preclinical data
  - Biocompatibility;
  - Sterilization and packaging;
  - Bench performance;
  - Software information;
  - Cybersecurity documentation;
  - ....
4. Clinical data (valid scientific evidence);
5. Labeling [+ OCE];
6. Manufacturing;
7. Postmarket study proposal;
8. ...

# Original PMA (FDA review team)



Engineer  
(BioMed)

Lead reviewer/  
Project manager

Statistician

Engineer  
(Mechanical)

Consumer safety  
officer

Epidemiologist

OCE Communication  
specialist



Scientist  
(Toxicologist)

OSEL Scientist

Medical officer

Scientist  
(Microbiologist)

# Original PMA (FDA advisory panel)

---



# Original PMA (FDA review process)



## Q-Sub Acceptance Checklist

Reviewer:  
Office/Division/Branch:  
Q Number:  
Device Name:  
Sponsor Name:  
RTA Recommendation:

1.	Has the type of cover letter provided in Q-Sub? Cf
	a. Pre
	b. Info
	c. Sub
	d. Ear
	(inc
	Det
	e. Stu

Note: this c

## Checklist for Acceptance Review for PMAs

(should be completed within 15 days of DCC receipt)

PMA Number: \_\_\_\_\_ Date Received: \_\_\_\_\_  
Device: \_\_\_\_\_ Procode: \_\_\_\_\_  
Company Name: \_\_\_\_\_  
Contact Name: \_\_\_\_\_  
Lead Reviewer: \_\_\_\_\_

## Checklist for Filing Review for PMAs

### Filing Assessment of Technical Elements – Clinical Studies

Check “Yes” if the information submitted is considered adequate to permit substantive review, “N/A” if it is not needed and “No” if it is not included.

Answers in the s

1. Is the product a constituent part of a combination product? Office Jurisdiction management. *P*
2. If the product is review by the Center with the appropriate CBER Office Jurisdiction management. *P*

A.	Consistency of study data sponsor’s study protocol recommendations from device-specific guidance
1.	Sample size/number the number of eval
2.	Follow-up duratio
3.	Follow-up evaluat
4.	Study Objectives
5.	Study Population/

## Worksheet for Benefit-Risk Determinations

Factor	Questions to Consider	Notes
<b>Assessment of Benefits of Devices</b>		
Type of benefit(s)	<ul style="list-style-type: none"> <li>- What primary endpoints or surrogate endpoints were evaluated?</li> <li>- What key secondary endpoints or surrogate endpoints were evaluated?</li> <li>- What value do patients place on the benefit?</li> </ul>	
Magnitude of the benefit(s)	<ul style="list-style-type: none"> <li>- For each primary and secondary endpoint or surrogate endpoints evaluated:                             <ul style="list-style-type: none"> <li>o What was the magnitude of each treatment effect?</li> </ul> </li> <li>- What scale is used to measure the benefit?                             <ul style="list-style-type: none"> <li>o How did the benefit rank on that scale?</li> </ul> </li> </ul>	
Probability of the patient	<ul style="list-style-type: none"> <li>- Was the study able to predict which</li> </ul>	



# Original PMA (FDA review timeline)

---

1. Pre-PMA Q-Sub (within 75 days)
2. Acceptance Review (within 15 days – FDA review clock starts here at receipt date if PMA is accepted and filed.)
3. Filing Review (within 45 days of receipt)
4. Substantive Review (within 90 days of receipt)
  - a. Major deficiency letter
    - i. Stops FDA review clock;
    - ii. Requires a Major Amendment with complete response to restart review clock.
  - b. Minor deficiency email/call
5. Day 100 Meeting (if applicant requested)
6. Advisory Panel Meeting (as needed)
  - a. Major deficiency letter, or
  - b. Minor deficiency email/call
7. Decision
  - a. No panel meeting (within 180 days)
  - b. Panel meeting (within 320 days)





# Original PMA (Panel Review)

---

## Advisory Panel Meeting (FDCA Sec. 515(c)(3)(B) & (f)(2)(B) and 21 CFR 814.44 and 814.116)

1. FDA convenes panel meeting when:
  - Device is first of its kind;
  - Concerns with performance, outcomes or study conducts; and/or
  - At the request of PMA applicant.
  
2. Panel constituents (Special Govn. Employees- SGEs 5 CFR 2640.103(a)(1) ):
  - Practitioners;
  - Biostatistician(s);
  - Technical expert(s);
  - Patient/consumer representative\*;
  - Industry representative\*;
  - FDA Office Director; and
  - FDA Designated Officer (DFO)

\*Not SGE



# Original PMA (Panel Review, cont'd)

---

## Advisory Panel Meeting (FDCA Sec. 515(c)(3)(B) & (f)(2)(B) and 21 CFR 814.44 and 814.116)

3. Pre-meeting, **applicant interacts with DFO** on Panel Pack:
  - Provide requested redacted/unredacted materials such as Executive Summary, PMA application, IDE protocol, Labeling, CRFs of certain subjects;
  - Comment on FDA Executive Summary (respond to FDA comments on applicant's Executive Summary); and,
  - Comment on FDA questions for the panel.
  
4. Panel Meeting process:
  1. Applicant presentation;
  2. FDA presentation;
  3. Open public hearing;
  4. Panel deliberations and answering FDA questions; and.
  5. Panel Voting.



## Original PMA (FDA review considerations)

---

### Safety and Effectiveness evaluation (FDCA Sec. 513(a)(2) and 21 CFR 860.7(b))

1. Intended patient population;
2. Intended conditions of use;
3. Weighing probable benefit against probable risk of injury or illness; and,
4. Reliability of the device (in CFR)



## Original PMA (FDA review considerations, cont'd)

---

PMA approval is based on (FDCA 515(d)(1)(A)(i) and 814.44 (d)(1)):

1. Reasonable assurance of device safety;
2. Reasonable assurance of device effectiveness;
3. Good manufacturing practices (FDCA 520(f) and CFR 820);
4. True and accurate labeling (per CFR 801 or 809); and
5. ...



## Original PMA (FDA review considerations, cont'd)

---

- Reasonable Assurance of **Safety**: when “it can be determined, based upon valid scientific evidence, that the **probable benefits ... outweigh any probable risks.**” (21 CFR 860.7(d)(1) )
- Reasonable Assurance of **Effectiveness**: when “it can be determined, based upon valid scientific evidence that **in a significant portion of the target population... the use of the device for its intended uses ... will provide clinically significant results.**” (21 CFR 860.7(e)(1) )



## Original PMA (FDA review considerations, cont'd)

---

### Other Approval Considerations for PMA Order:

- Extrapolate to **pediatric** population (FDCA 515A (b))
- Conditions of approval (CFR 814.82)
  - Restrictions of the sale, distribution or use (FDCA 515(d)(1)(B)(ii) and 520(e))
  - **Post-approval study**
  - **Post-market surveillance (522) study** (FDCA 522)
  - Tracking (FDCA 519(e) and CFR 821)
  - ...

# FDA Decision On PMA

---

- **Approval with conditions** (21CFR 814.44 (d))
- **Approvable pending...** (21CFR 814.44 (e))
  - QSR inspection
  - Agreement to approval conditions
  - ...
- **Not approvable** (21CFR 814.44 (f))
  - major deficiencies
- **Denial of approval** (21CFR 814.45)





## Polling Question on Post-Approval Changes

---

Post-approval, the PMA device, its manufacturing process, and/or its labeling can be modified, when...

Please choose one of the following:

1. FDA has provided approval for the change;
2. The company determines that the change is to enhance safety or can be evaluated by an already approved method (i.e., no prior FDA-approval is needed);
3. The change is to be eventually reported to FDA within the allotted time period (such as annually by the date of the original approval);
4. None of the above; or,
5. All of the above.





# **Post-Approval: PMA Submissions for Changes**

(FDCA 515(d)(6) and CFR814.39)

- **Amendment** (update to ownership, contact info.)
- **Special supplement** (certain labeling and manufacturing changes - 30 day)
- **30-day notice** (minor manufacturing change)
- **135-day supplement** (manufacturing change not qualified for 30-day notice)
- **180-day site change supplement**
- **Real time supplement** (minor change - 90 day)
- **180-day supplement** (significant change(s) that could affect safety or effectiveness)
  - **Real World Evidence** (significant labeling change(s))
- **Panel-Track supplement** (new indication – same clock as an Original PMA)

# PMA Annual Reports (FDCA 515(d)(6) and CFR814.39)

---



1. “Regular” annual report contains:
  - Changes FDA has approved or still under review;
  - **Changes implemented without FDA approval with justifications;**
  - Summary and bibliography of unpublished and published information on device, which were not previously submitted; and
  - Number of devices shipped or sold.
  
2. Post-approval study report

# Humanitarian Device Exemption Application (HDE)

---



# HDE



Number of Original HDEs*		
Review Office	FY 2020	FY 2021
OHT1		
OHT2		
OHT3		1
OHT4		
OHT5		
OHT6		2
OHT7		
<b>Total HDEs</b>	<b>0</b>	<b>3</b>

\*Number Approved per CDRH Database; number received is likely higher



## Original HDE (when applicable)

---

### Class III Humanitarian Use Device (HUD)

- HUD designation from FDA/Office of Orphan Products Development (OOPD)
- HUD criteria (FDCA 520(m) and CFR 814.102) :
  - Incidence rate of < 8000 patients per year; and,
  - No other legally marketed, “non-HUD” device for the intended use.



## Original HDE (contents - CFR 814.104)

---

1. OOPD's HUD Designation Letter with Annual Distribution Number (ADN);
2. Explanations for need of HDE;
3. Summary of safety and probable benefit (SSPB)
4. Indications;
5. Device Description;
6. Preclinical data [+ OSEL];
7. Clinical data (valid scientific evidence);
8. Labeling [+ OCE];
9. Manufacturing;
10. Postmarket study proposal;
11. ...

# Original HDE (FDA review teams)

---



Same as PMA:





# Original HDE (FDA review process)

---

1. **Filing Review** (within 30 days - CFR 814.112)
2. **Advisory Panel Meeting** (if needed)
  - a. **Major deficiency letter**

Major Amendment will reset FDA review clock back to 75 days.
  - b. **Minor deficiency email/calls**
3. **Decision** (within 75 days per FDCA 520(m)(2) and CFR 814.114)
  - Actual duration was ~8 mos. for one HDE and ~20 mos. for two HDEs approved in 2021.





## Original HDE (FDA review considerations)

---

HDE approval is based on (FDCA 520(m) and CFR 814.118):

1. Reasonable assurance of device safety;
2. Probable benefit to health;
3. Good manufacturing practices (FDCA 520(f) and CFR 820);
4. True and accurate labeling (per CFR 801 or 809);
5. ...



## Original HDE (FDA review considerations, cont'd)

---

### Other Approval Considerations for HDE Order (same as PMA):

- Extrapolate to pediatric population (FDCA 515A (b))
- Conditions of approval (CFR 814.116 (c))
  - Restrictions of the sale, distribution or use (FDCA 515(d)(1)(B)(ii) and 520(e))
  - Post-approval study
  - Post-market surveillance study (FDCA 522)
  - Tracking (FDCA 519(e) and CFR 821)
  - ...

# FDA Decisions On HDE (same as PMA)

---

- Approval with conditions (21CFR 814.116 (b))
- Approvable pending... (21CFR 814.116 (c))
  - QSR inspection
  - Agreement to approval conditions
  - ...
- Not approvable (21CFR 814.116 (d))
  - major deficiencies
- Denial of approval (21CFR 814.118)



# HDE Submissions for Post-Approval Changes (CFR814.108)

---



- 30 day notice (minor manufacturing change)
  - 75 day supplement (manufacturing change not qualified for 30 day notice)
  - 75 day supplement (design or labeling change)
- **New indication** for an HDE device requires new HUD and HDE.



## HDE Annual Reports (CFR814.126(b))

---

1. “Regular” annual report contains:
  - Updated annual incidence reassessment (AIR) to justify HUD designation;
  - Updated explanations of HDE eligibility;
  - Updated analysis of risks and benefits;
  - Amount charged and justifications if over \$250;
  - Number of devices shipped or sold;
  - Clinical experience with device; and
  - Summary of changes implemented following approved supplements.
  
2. Post-approval study report



## HDE (postmarket considerations )

---

### Restrictions on HDE-approved device (FDCA 520(m)):

- No profit on sales,
  - Unless intended for pediatric patients (*subject to annual review by Pediatric Advisory Panel*);
- IRB approval and monitoring (FDCA 520 (m) (4) and CFR814.124); and,
- Not exceeding ADN.



HDE (postmarket considerations, cont'd )

---

**Withdrawal of HDE Approval Order** when a comparable device is PMA- approved for the same intended use (FDCA 520(m)(2)(B))

# Breakthrough and Safer Designations

—





# Breakthrough Device

(FDA 2018 Guidance)



**Breakthrough Device Designation** (FDCA 515(c) – to be  
— requested in a Q-Sub)

1. Provide for **more effective** treatment or diagnosis of life-threatening or irreversibly debilitating disease or conditions; and
2. one of the following:
  - a. breakthrough technology;
  - b. no approved alternative;
  - c. offers significant [clinically meaningful] advantages over existing approved or cleared alternatives; or
  - d. availability is in the best interest of patients.



## Breakthrough PMA (how it's reviewed?)

---

PMA Process but Approval could be based on:

1. Intermediate or Surrogate endpoints
2. Smaller sample in premarket phase (with additional data in postmarket phase)
3. Less manufacturing information in PMA (if company has a good QSR track record)
4. No pre-approval inspection of facility (with good inspectional history)



# **Safer Technologies Program (STeP)**

(FDA 2021 Guidance)



## STeP Designation (to be requested in a Q-Sub):

1. Not eligible for Breakthrough Designation due to the less serious nature of the disease or condition; and,
2. **Safer** (“significantly improve benefit-risk”) treatment or diagnostic in one of the following areas:
  - a. Reducing occurrence of a known serious adverse event;
  - b. Reducing occurrence of a known failure mode;
  - c. Reducing occurrence of a known use-related hazard or use error; or
  - d. Improving the safety of another device or intervention.



## STeP PMA (how it's reviewed?)

---

PMA Process but Approval could be based on:

1. Less manufacturing information in PMA (if company has a good QSR track record)
2. No pre-approval inspection of facility (with good inspectional history)



## Polling Question on Breakthrough and STeP

---

The applicant should submit the Q-Sub to request Breakthrough or STeP Designation...

Please choose one of the following:

1. Once preliminary results are available (during the IDE stage or sooner);
2. As a part of the Pre-PMA Q-Sub; or,
3. At any time while FDA is reviewing the PMA submission.

## Key Takeaways

- Each pathway has its own set of requirements.
- Write submission to address each requirement.
- Use language from FDA's checklist or guidance, if available, to help form the arguments and justifications.





# Key Resources

---

- FDA 'Device Advice' on PMA, <https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PreMarketSubmissions/PreMarketApprovalPMA/default.htm>
- FDA 'Device Advice' on HDE, <https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/PreMarketSubmissions/HumanitarianDeviceExemption/default.htm>
- Consideration of Uncertainty in Making Benefit-Risk Determinations in Medical Device Premarket Approvals, De Novo Classifications, and Humanitarian Device Exemptions, final guidance issued on August 2019, <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM619220.pdf>
- Modifications to Devices Subject to Premarket Approval (PMA) - The PMA Supplement Decision-Making Process, <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/modifications-devices-subject-premarket-approval-pma-pma-supplement-decision-making-process>
- Breakthrough Devices Program, <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/breakthrough-devices-program>
- Safer Technologies Program (STeP), <https://www.fda.gov/medical-devices/how-study-and-market-your-device/safer-technologies-program-step-medical-devices>

# Thanks!

---



**Quynh Hoang**  
Senior Regulatory Consultant  
*FDA and Life Sciences*

qhoang@kslaw.com  
+1 202 626 2939